

EVOLUTION

PATIENT POSITIONERS

Návod k použití Polohovače v leže na zádech

HEAD – PAGE 3					
Covers		Supine Head Positioners			
REF	CSHP12	REF	SHP12	malý	Styl 1
REF	CSHP14	REF	SHP14	velký	Styl 1
REF	CSHP23	REF	SHP23	střední	Styl 2
REF	CSHP24	REF	SHP24	velký	Styl 2

ARM – PAGE 4		
Supine Arm Positioner		
REF	SAP14	velký
REF	SAP13	střední
Strap And Retainer Kit		
REF	SRK14	velký

TORSO - PAGE 5			
Supine Torso Positioners			
REF	STPL14	Left	velký
REF	STPR14	Right	velký

KNEE AND FOOT – PAGE 6			
Supine Foot Positioners			
REF	SFPL14	velký	Left
REF	SFPR14	velký	Right
REF	SFPP13	střední	Universal
Supine Knee Positioners			
REF	SKPP14	velký	Universal
REF	SKPP13	střední	Universal

PRE-OP GENERAL	
Page	Wording
2	23 INSPECT BEFORE USE
	24 USE ON DRY SURFACE
	25 KEEP DRY
	28 UNSUITABLE USE 1. DOUBT 2.OUT
	210 DO NOT OVER STRAIN LIMBS
	211 AVOID EXAGGERATED STRESSES OF POSITIONERS

HEAD	
Page	Wording
3	216 DO NOT USE DAMAGED COVERS
	217 PRODUCT INFORMATION
	222 KEEP DRY
	224 DO NOT COVER APERTURE
	229 PRODUCT MISUSE

Datum vydání: 1st July 2

Czech

ARM	
Page	Wording
4	231 SUPINE ARM POSITIONER
	232 PATIENT HEIGHT & WEIGHT LIMITS
	233 PRODUCT INFORMATION
	235 ELBOW
	236 STRAP ARM BOARD
	238 UNIVERSAL
	239 PRODUCT MISUSE
	240 FASTEN

TORSO	
Page	Wording
5	242 SUPINE TORSO POSITIONER Left/Right
	244 PATIENT HEIGHT & WEIGHT LIMITS
	245 PRODUCT INFORMATION
	249 PRODUCT MISUSE

KNEE & FOOT	
Page	Wording
6	251 SUPINE KNEE & FOOT POSITIONERS Left/Right
	253 PRODUCT INFORMATION
	254 USE TOGETHER
	258 PRODUCT MISUSE
	261 UNIVERSAL

OPERATION	
Page	Wording
7	263 DO NOT USE AT SURGICAL SITE
	264 AVOID FRICTION
	265 AVOID POOLING OF FLUIDS
	266 OEDEMA MONITORING
	267 AVOID CONDITIONS OF ELECTRO-STATIC DISCHARGE
	268 MONITOR COVER

POST-OP	
Page	Wording
8	269 HOSPITAL EQUIPMENT CLEANING/DISINFECTION PROCEDURE
	270 ACTIVE INGREDIENTS
	<input type="checkbox"/> SODIUM HYPOCHLORIDE <input type="checkbox"/> SODIUM CHLORIDE <input type="checkbox"/> PROPAN-1-OL (ALCOHOL WIPES) <input type="checkbox"/> SODIUM DICHLOROISOCYANURATE
	RECOMMENDED DISINFECTANT TO USE
	271 STANDARD DETERGENT
	272 DISINFECTANT SOAK
	273 GAS STERILISATION
	274 INSPECT AFTER USE
	275 DISINFECTANT STAINING OK
	276 END OF PRODUCT LIFE
	DEFORMED CRACKED

	TORN
	② 78 NON-CONDUCTIVE
	② 79 PLAIN FILM X-RAY, CT SCAN
	② 82 STORAGE
	② 85 KEEP STOCK

Jakýkoli závažný incident, ke kterému došlo v souvislosti s tímto zařízením, by měl být nahlášen výrobcí a příslušnému orgánu členského státu, ve kterém sídlí uživatel.